

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/31/2025

OPERATOR NAME:
ANDREW J MARANVILLE
Permit #: 53581
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:55
INTERNAL STANDARD	VERIFIED	00:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59

