

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/01/2025

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:20
INTERNAL STANDARD	VERIFIED	05:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:20
BLANK TEST	0.00	05:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24

