

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/07/2025

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52

