

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/21/2025

OPERATOR NAME:
ANDREW J MARANVILLE
Permit #: 53581
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:17
INTERNAL STANDARD	VERIFIED	05:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:17
BLANK TEST	0.00	05:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:19
BLANK TEST	0.00	05:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:20
BLANK TEST	0.00	05:21

