

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/01/2025

OPERATOR NAME:
DANELLE M BELL
Permit #: 53551
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:00
INTERNAL STANDARD	VERIFIED	05:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:02
BLANK TEST	0.00	05:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:04

