

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 06/09/2025

OPERATOR NAME:  
REESE P WHEATON  
Permit #: 53611  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:29
INTERNAL STANDARD	VERIFIED	02:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32

