

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/16/2025

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:08 |
| INTERNAL STANDARD | VERIFIED | 00:08 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:09 |
| BLANK TEST | 0.00 | 00:10 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:10 |
| BLANK TEST | 0.00 | 00:11 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:11 |
| BLANK TEST | 0.00 | 00:12 |

