

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 06/23/2025

OPERATOR NAME:  
ANDREW J MARANVILLE  
Permit #: 53581  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:21
INTERNAL STANDARD	VERIFIED	00:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:21
BLANK TEST	0.00	00:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:23
BLANK TEST	0.00	00:24

