

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/30/2025

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:53
INTERNAL STANDARD	VERIFIED	01:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:56
BLANK TEST	0.00	01:57

