

SOLUTION CHANGE

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 07/01/2025

OPERATOR NAME:  
 MICHAELA R ORTALE  
 Permit #: 52289  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:17
INTERNAL STANDARD	VERIFIED	06:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:17
BLANK TEST	0.00	06:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:19
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:20
BLANK TEST	0.00	06:21

