

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/14/2025

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:34
INTERNAL STANDARD	VERIFIED	04:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:34
BLANK TEST	0.00	04:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:35
BLANK TEST	0.00	04:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:36
BLANK TEST	0.00	04:37

