

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/21/2025

OPERATOR NAME:
ANDREW J MARANVILLE
Permit #: 53581
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21

