

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/29/2025

OPERATOR NAME:
ANDREW J MARANVILLE
Permit #: 53581
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53

