

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 10/01/2025

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:06
INTERNAL STANDARD	VERIFIED	06:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:07
BLANK TEST	0.00	06:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:09
BLANK TEST	0.00	06:10

