

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/06/2025

OPERATOR NAME:  
ANDREW J MARANVILLE  
Permit #: 53581  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:40
INTERNAL STANDARD	VERIFIED	01:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43

