

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 11/03/2025

OPERATOR NAME:  
REESE P WHEATON  
Permit #: 53611  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43

