

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/17/2025

OPERATOR NAME:
RRSE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:04
INTERNAL STANDARD	VERIFIED	02:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07

