

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/01/2025

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:34
INTERNAL STANDARD	VERIFIED	01:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:37
BLANK TEST	0.00	01:38

