

SUPERVISOR TEST

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 12/08/2025

OPERATOR NAME:  
 ANDREW J MARANVILLE  
 Permit #: 53581  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 25330

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 01:22 |
| INTERNAL STANDARD       | VERIFIED | 01:22 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 01:23 |
| BLANK TEST              | 0.00     | 01:24 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 01:24 |
| BLANK TEST              | 0.00     | 01:25 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 01:25 |
| BLANK TEST              | 0.00     | 01:26 |

