

SUPERVISOR TEST

-----  
GLENNS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 12/15/2025

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GLENNS FALLS CITY PD  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:32
INTERNAL STANDARD	VERIFIED	14:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36

