

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/22/2025

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:11
INTERNAL STANDARD	VERIFIED	05:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:12
BLANK TEST	0.00	05:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:13
BLANK TEST	0.00	05:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:14
BLANK TEST	0.00	05:15

