

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 12/29/2025

OPERATOR NAME:  
REESE P WHEATON  
Permit #: 53611  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:20
INTERNAL STANDARD	VERIFIED	02:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24

