

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 01/01/2026

OPERATOR NAME:
 DANELLE M BELL
 Permit #: 53551
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35

