

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 01/26/2026

OPERATOR NAME:  
HUNTER L BARTON  
Permit #: 49353  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:34
INTERNAL STANDARD	VERIFIED	02:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38

