

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/01/2026

OPERATOR NAME:  
REESE WHEATON  
Permit #: 53611  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:08
INTERNAL STANDARD	VERIFIED	02:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:11
BLANK TEST	0.00	02:12

