

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/09/2026

OPERATOR NAME:
REESE WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:35
INTERNAL STANDARD	VERIFIED	04:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:36
BLANK TEST	0.00	04:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:37
BLANK TEST	0.00	04:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:38
BLANK TEST	0.00	04:39

