

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/16/2026

OPERATOR NAME:  
REESE WHEATON  
Permit #: 53611  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:28
INTERNAL STANDARD	VERIFIED	04:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:28
BLANK TEST	0.00	04:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:29
BLANK TEST	0.00	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:32

