

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/23/2026

OPERATOR NAME:  
HUNTER S MCKENNA  
Permit #: 52925  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:43
INTERNAL STANDARD	VERIFIED	04:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:44
BLANK TEST	0.00	04:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:46
BLANK TEST	0.00	04:47

