

SOLUTION CHANGE

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 03/01/2026

OPERATOR NAME:  
 MICHAELA R ORTALE  
 Permit #: 52289  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:22
INTERNAL STANDARD	VERIFIED	04:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:23
BLANK TEST	0.00	04:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26

