

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/03/2026

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:04
INTERNAL STANDARD	VERIFIED	10:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:04
BLANK TEST	0.00	10:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:05
BLANK TEST	0.00	10:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:06
BLANK TEST	0.00	10:07

