

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/23/2026

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:20
INTERNAL STANDARD	VERIFIED	06:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:21
BLANK TEST	0.00	06:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:23
BLANK TEST	0.00	06:24

