


**BREATH SCREENING DEVICE  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Glens Falls City Police Department**  
**Manufacturer: INTOXIMETERS INC      Model: ALCO-SENSOR FST**  
**Serial Number: 244250**  
**Date: July 09, 2020**

**Reference Solution Lot Number: 20040**  
**Reference Solution Standard: 0.100 % at 34.0 degrees centigrade**

<b>Test No</b>	<b>Test Result</b>	<b>Simulator Temperature</b>
1	0.098 %	34.02 degrees centigrade
2	0.098 %	34.03 degrees centigrade
3	0.097 %	34.02 degrees centigrade
4	0.097 %	34.03 degrees centigrade

  
\_\_\_\_\_  
**George Coffin**  
**Highway Safety Equipment Technician**  
**Office of Public Safety**

**BREATH SCREENING DEVICE  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Glens Falls City Police Department**  
**Manufacturer: INTOXIMETERS INC**  
**Model: ALCO-SENSOR FST**  
**Serial Number: 244250**  
**Date: July 09, 2020**

**1. General Appearance and condition of this device -----: OK**

**2. Check battery indicators -----: OK**

**3. Check zero flashing -----: OK**

**4. Record any repairs made to this instrument and list services provided: VERIFIED  
CALIBRATION**



  
\_\_\_\_\_  
**George Coffin**  
**Highway Safety Equipment Technician**  
**Office of Public Safety**

