



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP4259, performed on May 09, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 05/20/2022 07:01 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department
Manufacturer: GUTH LABORATORIES
Model: 12V500
Serial Number: MP4259
Date: May 09, 2022

Breath Simulator Temperature
34.04 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4259, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink that reads "George Coffin".

Digitally signed under ESRA
by George Coffin
on 05/09/2022 01:49 PM

George Coffin
Highway Safety Equipment Technician
Office of Public Safety

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP4259

Date: May 09, 2022

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

**5. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



**Division of Criminal
Justice Services**

A handwritten signature in cursive script that reads "George Coffin".

Digitally signed under ESRA
by George Coffin
on 05/09/2022 01:49 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**

NIST Traceable
Calibration Report



Reference Number: **1405165**
 PO Number: **SCALUCCIO061421**

New York State Division Criminal Justice Services
 80 South Swan Street
 Albany, NY 12210 United States

Manufacturer: Ertco
Model Number: 4400
Description: Temperature, Dig thermometer w/probe
Asset Number: 304011
Serial Number: 304011
Procedure: DS Eutechnics 4400-EP642

Calibration Date: 06/17/2021
Calibration Due Date: 06/17/2022
Condition As Found: In Tolerance
Condition As Left: In Tolerance, No adjustment

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Utilized

Asset No.	Manufacturer	Model No.	Description	Cal. Date	Due Date
CP144317	Fluke Corporation	1595A	Thermometer, Readout, Precision Thermometer	01/28/2020	02/28/2022
CP32032	Hart Scientific	5628	Temperature, Probe	10/19/2020	04/30/2022
CP32034	Hart Scientific	5628	Temperature, Probe	01/22/2020	10/31/2021
CP32042	Hart Scientific	5680	Temperature, Probe, SPRT	03/31/2021	04/30/2022
CP53595	Hart Scientific	5628	Temperature, Probe	01/28/2020	02/28/2022

Calibration Data

FUNCTION TESTED	Nominal Value	As Found	Out of Tol	As Left	Out of Tol	CALIBRATION TOLERANCE
Temperature Accuracy	-15.000 °C	-14.97		Same		-15.050 to -14.950 °C [EMU 0.013 °C][TUR 3.8:1]
	0.000 °C	0.01		Same		-0.030 to 0.030 °C [EMU 0.012 °C][TUR 2.5:1]
	75.000 °C	75.01		Same		74.970 to 75.030 °C [EMU 0.015 °C][TUR 2.0:1]
	125.000 °C	124.99		Same		124.950 to 125.050 °C [EMU 0.015 °C][TUR 3.4:1]

Temperature: 21° C
Humidity: 49% RH
Rpt. No.: 1655081

Calibration Performed By:				Quality Reviewer:	
Name	ID #	Title	Phone	Name	Date
Sittinger, Jennifer	330	Metrologist	847-327-5317	Ziegler, Jeff	06/18/2021

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002-94, and ISO 17025:2017. Guard Banding, if reported on this certificate, is applied at a Z-factor of 30% for test points with a test uncertainty ratio (TUR) below 4:1. In Tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

* GRL
5/9/22


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency GLENS FALLS POLICE DEPARTMENT		Today's Date 05/03/2022
Address 42 RIDGE ST		City, State, ZIP GLENS FALLS, NY 12801
Contact Person LT MATTISON		Contact Telephone 518-761-3840 X 3782
Contact Email CMATTISON@GLENSFALLSPD.COM		Best Way to Reach You During Appointment Period PHONE/EMAIL

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS <u>DMT</u> GUTH Applied Concepts (Stalker) Decatur (Genesis) Kustom (Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number SIMULATOR FOR DMT		Instrument Serial Number MP4259
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction YEARLY CERTIFICATION-SERVICE APPOINTMENT SCHEDULED FOR 05/09/2022		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT <u>SIM</u> ⁺ Tubes Keyboard Other:		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature  5375		Date 05/03/2022

DCJS USE ONLY

Export °C	Digital °C	Thermometer: <u>304011</u>
<u>34.03</u>	<u>34.04</u>	Calibration Due: <u>6/17/22</u>
<u>34.03</u>	<u>34.04</u>	Air Leak Test: <u>OK</u> Propeller Rotation: <u>OK</u>
<u>34.03</u>	<u>34.04</u>	