

SUPERVISOR TEST

LAKE GEORGE  
 PARK COMMISSION  
 DATAMASTER dmt: 300391

Date: 05/11/2021

OPERATOR NAME:  
 SHANE L ROSS  
 Permit #: 12663  
 Operator Agency/Dept: LAKE GEORGE PK COMM  
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:07
INTERNAL STANDARD	VERIFIED	13:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:10
BLANK TEST	0.00	13:11

