

SUPERVISOR TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 05/17/2021

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:48
INTERNAL STANDARD	VERIFIED	08:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:51
BLANK TEST	0.00	08:52

