

SUPERVISOR TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 05/31/2021

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:19
INTERNAL STANDARD	VERIFIED	11:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:20
BLANK TEST	0.00	11:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:21
BLANK TEST	0.00	11:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:22
BLANK TEST	0.00	11:23

