

SUPERVISOR TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 06/07/2021

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:34
INTERNAL STANDARD	VERIFIED	09:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:34
BLANK TEST	0.00	09:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38

