

SUPERVISOR TEST

LAKE GEORGE  
 PARK COMMISSION  
 DATAMASTER dmt: 300391

Date: 06/14/2021

OPERATOR NAME:  
 SHANE L ROSS  
 Permit #: 12663  
 Operator Agency/Dept: LAKE GEORGE PK COMM  
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:58
INTERNAL STANDARD	VERIFIED	10:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:01
BLANK TEST	0.00	11:02

