

SUPERVISOR TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 06/28/2021

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:49
INTERNAL STANDARD	VERIFIED	10:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53

