

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: LGPC

Instrument Serial #: 300391

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/11/21	21020	34.00	34.00	34.00	MP5081	109	110	110	Pass	In Service
5/17/21	21020	34.00	34.00	34.00	MP5081	109	109	109	Pass	
5/24/21	21020	34.01	34.00	34.00	MP5081	109	109	109	Pass	
5/31/21	21020	34.00	34.01	34.00	MP5081	109	109	109	Pass	
6/5/21	21100	34.00	34.00	34.01	MP5081	109	109	109	Pass	YES
6/7/21	21100	34.00	34.00	34.00	MP5081	109	107	109	Pass	
6/11/21	21100	34.00	34.01	34.00	MP5081	109	109	109	Pass	
6/21/21	21100	34.01	34.01	34.01	MP5081	109	109	109	Pass	
6/28/21	21100	34.01	34.00	34.01	MP5081	109	109	109	Pass	
7/3/21	21100	34.00	34.00	34.00	MP5081	109	109	109	Pass	YES
7/5/21	21100	34.00	34.01	34.01	MP5081	109	109	109	Pass	
7/13/21	21100	34.01	34.01	34.01	MP5081	109	109	109	Pass	

* - Please enter a ✓ to denote simulator solution change taking place