

SOLUTION CHANGE

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 05/01/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:45
INTERNAL STANDARD	VERIFIED	12:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:46
BLANK TEST	0.00	12:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:47
BLANK TEST	0.00	12:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:48
BLANK TEST	0.00	12:49

