

SUPERVISOR TEST

LAKE GEORGE  
PARK COMMISSION  
DATAMASTER dmt: 300391

Date: 05/09/2022

OPERATOR NAME:  
SHANE L ROSS  
Permit #: 12663  
Operator Agency/Dept: LAKE GEORGE PK COMM  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:15
INTERNAL STANDARD	VERIFIED	08:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:19

