

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 05/24/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:22
INTERNAL STANDARD	VERIFIED	13:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:22
BLANK TEST	0.00	13:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:24
BLANK TEST	0.00	13:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:25
BLANK TEST	0.00	13:26

