

SOLUTION CHANGE

LAKE GEORGE  
PARK COMMISSION  
DATAMASTER dmt: 300391

Date: 06/06/2022

OPERATOR NAME:  
SHANE L ROSS  
Permit #: 12663  
Operator Agency/Dept: LAKE GEORGE PK COMM  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:18
INTERNAL STANDARD	VERIFIED	10:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22

