

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 06/20/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:44
INTERNAL STANDARD	VERIFIED	10:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:45
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:47
BLANK TEST	0.00	10:48

