

SOLUTION CHANGE

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 07/03/2022

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:48
INTERNAL STANDARD	VERIFIED	13:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:51
BLANK TEST	0.00	13:51

