

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 07/11/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:55
INTERNAL STANDARD	VERIFIED	09:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59

