

SOLUTION CHANGE

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 07/22/2022

OPERATOR NAME:
ANGELO P PACCIONE
Permit #: 47655
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:34
INTERNAL STANDARD	VERIFIED	17:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:35
BLANK TEST	0.00	17:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:36
BLANK TEST	0.00	17:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:37
BLANK TEST	0.00	17:38

