

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 08/08/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:07
INTERNAL STANDARD	VERIFIED	10:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:07
BLANK TEST	0.00	10:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:10
BLANK TEST	0.00	10:11

