

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 08/15/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:52
INTERNAL STANDARD	VERIFIED	11:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:55
BLANK TEST	0.00	11:56

